



Product Service

CERTIFICATE

No. Z1 17 08 89832 006

Holder of Certificate: **SUZHOU ANTAI AIRTECH CO., LTD**No.2 Weixin Road, SIP
215122 Suzhou, Jiangsu
PEOPLE'S REPUBLIC OF CHINA

Production Facility(ies): 91890

Certification Mark:

Product: **Laboratory Equipment
Bio. Safety Cabinet**Model(s): **BSC-1003IIA2, BSC-1003IIA2-S, BSC-1003IIA2-E
BSC-1003IIA2-SE, BSC-1303IIA2, BSC-1303IIA2-S
BSC-1303IIA2-E, BSC-1303IIA2-SE, BSC-1603IIA2
BSC-1603IIA2-S, BSC-1603IIA2-E, BSC-1603IIA2-SE**

Parameters:	Rated input voltage:	220-240VAC
	Rated input frequency:	50/60Hz
	Max. power:	1.5kVA(BSC-1003IIA2 BSC-1003IIA2-S, BSC-1003IIA2-E BSC-1003IIA2-SE, BSC-1303IIA2 BSC-1303IIA2-S, BSC-1303IIA2-E BSC-1303IIA2-SE) 1.8kVA(BSC-1603IIA2 BSC-1603IIA2-S, BSC-1603IIA2-E BSC-1603IIA2-SE)
	Class of MSC:	II
	Electrical protection class:	I

Tested according to: EN 12469:2000
EN 61010-1:2010

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.: 884001506302
Valid until: 2022-08-21Date, 2017-09-06
Page 1 of 1

(Bin Li)
