



Product Service

CERTIFICATE

No. Z1 17 08 89832 006

Holder of Certificate: **SUZHOU ANTAI AIRTECH CO., LTD**

No.2 Weixin Road, SIP
215122 Suzhou, Jiangsu
PEOPLE'S REPUBLIC OF CHINA

**Production
Facility(ies):**

91890

Certification Mark:**Product:**

**Laboratory Equipment
Bio. Safety Cabinet**

Model(s):

**BSC-1003IIA2, BSC-1003IIA2-S, BSC-1003IIA2-E
BSC-1003IIA2-SE, BSC-1303IIA2, BSC-1303IIA2-S
BSC-1303IIA2-E, BSC-1303IIA2-SE, BSC-1603IIA2
BSC-1603IIA2-S, BSC-1603IIA2-E, BSC-1603IIA2-SE**

Parameters:

Rated input voltage:	220-240VAC
Rated input frequency:	50/60Hz
Max. power:	1.5kVA(BSC-1003IIA2 BSC-1003IIA2-S, BSC-1003IIA2-E BSC-1003IIA2-SE, BSC-1303IIA2 BSC-1303IIA2-S, BSC-1303IIA2-E BSC-1303IIA2-SE) 1.8kVA(BSC-1603IIA2 BSC-1603IIA2-S, BSC-1603IIA2-E BSC-1603IIA2-SE)
Class of MSC:	II
Electrical protection class:	I


Tested according to: EN 12469:2000
EN 61010-1:2010

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

Test report no.: 884001506302

Valid until: 2022-08-21

Date, 2017-09-06


(Bin Li)

Page 1 of 1





Product Service

Attestation of Conformity

No. E8A 17 07 89832 001

Holder of Certificate: SUZHOU ANTAI AIRTECH CO., LTD

 No.2 Weixin Road, SIP
 215122 Suzhou, Jiangsu
 PEOPLE'S REPUBLIC OF CHINA

Name of Object: Laboratory Equipment
 Bio. Safety Cabinet

Model(s): BSC-1003IIA2, BSC-1303IIA2, BSC-1603IIA2
 BSC-1003IIA2-S, BSC-1303IIA2-S, BSC-1603IIA2-S
 BSC-1003IIA2-E, BSC-1303IIA2-E, BSC-1603IIA2-E
 BSC-1003IIA2-SE, BSC-1303IIA2-SE, BSC-1603IIA2-SE

Description of Object:

Rated voltage:	220-240VAC, 50/60Hz
Max. power consumption:	1.5kW (BSC-1003IIA2, BSC-1303IIA2 BSC-1003IIA2-S, BSC-1303IIA2-S BSC-1003IIA2-E, BSC-1303IIA2-E BSC-1003IIA2-SE, BSC-1303IIA2-SE) 1.8kW (BSC-1603IIA2, BSC-1603IIA2-S BSC-1603IIA2-E, BSC-1603IIA2-SE)
Rated input voltage:	220-240VAC, 50/60Hz
Class of MSC:	II
Protection class:	I

Tested according to: EN 61326-1:2013

This Attestation of Conformity is issued on a voluntary basis according to the Directive 2014/30/EU relating to electromagnetic compatibility. It confirms that the listed apparatus complies with all essential requirements of the directive and is based on the technical specifications applicable at the time of issuance. It refers only to the particular sample submitted for testing and certification. See also notes overleaf.

Test report no.: 888001506402

Date, 2017-08-10

(Jing Zhang)



After preparation of the necessary technical documentation as well as the EU declaration of conformity the required CE marking can be affixed on the product. That declaration of conformity is issued under the sole responsibility of the manufacturer. Other relevant EU-directives have to be observed.

Page 1 of 1



Product Service

Attestation of Conformity

No. N8A 17 08 89832 005

Holder of Certificate: SUZHOU ANTAI AIRTECH CO., LTD

 No.2 Weixin Road, SIP
 215122 Suzhou, Jiangsu
 PEOPLE'S REPUBLIC OF CHINA

Product: Laboratory Equipment
 Bio. Safety Cabinet

Model(s): BSC-1003IIA2, BSC-1003IIA2-S, BSC-1003IIA2-E
 BSC-1003IIA2-SE, BSC-1303IIA2, BSC-1303IIA2-S
 BSC-1303IIA2-E, BSC-1303IIA2-SE, BSC-1603IIA2
 BSC-1603IIA2-S, BSC-1603IIA2-E, BSC-1603IIA2-SE

Parameters:

Rated input voltage:	220-240VAC
Rated input frequency:	50/60Hz
Max. power:	1.5kVA(BSC-1003IIA2 BSC-1003IIA2-S, BSC-1003IIA2-E BSC-1003IIA2-SE, BSC-1303IIA2 BSC-1303IIA2-S, BSC-1303IIA2-E BSC-1303IIA2-SE) 1.8kVA(BSC-1603IIA2 BSC-1603IIA2-S, BSC-1603IIA2-E BSC-1603IIA2-SE)
Class of MSC:	II
Electrical protection class:	I

Tested according to: EN 12469:2000
 EN 61010-1:2010

This Attestation of Conformity is issued on a voluntary basis according to the Low Voltage Directive 2014/35/EU relating to electrical equipment designed for use within certain voltage limits. It confirms that the listed equipment complies with the principal protection requirements of the directive and is based on the technical specifications applicable at the time of issuance. It refers only to the particular sample submitted for testing and certification. See also notes overleaf.

Test report no.: 884001506302

Date, 2017-09-06

(Bin Li)


After preparation of the necessary technical documentation as well as the EU declaration of conformity the required CE marking can be affixed on the product. The declaration of conformity is issued under the sole responsibility of the manufacturer. Other relevant EU-directives have to be observed.

Page 1 of 1