



Product Service

Attestation of Conformity

No. N8A 089832 0017 Rev. 00

Holder of Attestation: SUZHOU ANTAI AIRTECH CO., LTDNo.2 Weixin Road, SIP
215122 Suzhou, Jiangsu
PEOPLE'S REPUBLIC OF CHINA**Product:****Laboratory Equipment
CYTOTOXIC SAFETY CABINET****Model(s):****BSC-1004IIA2-c, BSC-1304IIA2-c, BSC-1604IIA2-c****Parameters:**Rated voltage 220-240V~
Rated frequency 50/60Hz
Rated power 2.15KVA (BSC-1004IIA2-c),
2.25KVA(BSC-1304IIA2-c),
2.35KVA (BSC-1604IIA2-c)Protection class I
Degree of protection IP20

EN 61010-1:2010/A1:2019

Tested**according to:**

This Attestation of Conformity is issued on a voluntary basis according to the Low Voltage Directive 2014/35/EU relating to electrical equipment designed for use within certain voltage limits. It confirms that the listed equipment complies with the principal protection requirements of the directive and is based on the technical specifications applicable at the time of issuance. It refers only to the particular sample submitted for conformity assessment. For details see: www.tuvsud.com/ps-cert

Test report no.:

4842023609400

Date,

2024-04-18

(Bin Li)

Page 1 of 1

This Attestation does not replace the regulatory EU Declaration of Conformity (DoC) and does not allow for CE marking. After preparation of the necessary documentation and establishing compliance to requirements of all applicable directives, the manufacturer may sign a DoC and apply the CE marking. The DoC is issued under the sole responsibility of the manufacturer.