



Product Service

CERTIFICATE

No. Z1 089832 0006 Rev. 01

Holder of Certificate: SUZHOU ANTAI AIRTECH CO., LTD

No.2 Weixin Road, SIP
215122 Suzhou, Jiangsu
PEOPLE'S REPUBLIC OF CHINA

Certification Mark:




Product: Laboratory Equipment
Bio. Safety Cabinet

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition, the certification holder must not transfer the certificate to third parties. This certificate is valid until the listed date, unless it is cancelled earlier. All applicable requirements of the testing and certification regulations of TÜV SÜD Group have to be complied. For details see: www.tuvsud.com/ps-cert

Test report no.: 4842022108000

Valid until: 2027-08-09

Date, 2022-08-10


(Bin Li)



CERTIFICATE

No. Z1 089832 0006 Rev. 01

Model(s): BSC-1003IIA2, BSC-1003IIA2-S,
 BSC-1003IIA2-E, BSC-1003IIA2-SE,
 BSC-1303IIA2, BSC-1303IIA2-S,
 BSC-1303IIA2-E, BSC-1303IIA2-SE,
 BSC-1603IIA2, BSC-1603IIA2-S,
 BSC-1603IIA2-E, BSC-1603IIA2-SE

Parameters:
 Rated voltage: 220-240V~
 Rated frequency: 50/60Hz
 Rated power: 1.5kVA (BSC-1003IIA2,BSC-1003IIA2-S,
 BSC-1003IIA2-E, BSC-1003IIA2-SE,
 BSC-1303IIA2, BSC-1303IIA2-S,
 BSC-1303IIA2-E, BSC-1303IIA2-SE);
 1.8kVA (BSC-1603IIA2, BSC-1603IIA2-S,
 BSC-1603IIA2-E, BSC-1603IIA2-SE)
 Protection class: I
 Degree of protection: IP20

Tested according to: EN 61010-1:2010/A1:2019
 EN 12469:2000