



Attestation of Conformity

No. N8A 123678 0003 Rev. 00

Holder of Attestation: **Qingdao Antech Scientific Co., Ltd.**

No.182 Haier Road, Laoshan District
266100 Qingdao, Shandong
PEOPLE'S REPUBLIC OF CHINA

Product: **Laboratory equipment
(Biological Safety Cabinet)**

This Attestation of Conformity is issued on a voluntary basis according to the Low Voltage Directive 2014/35/EU relating to electrical equipment designed for use within certain voltage limits. It confirms that the listed equipment complies with the principal protection requirements of the directive and is based on the technical specifications applicable at the time of issuance. It refers only to the particular sample submitted for conformity assessment. For details see: www.tuvsud.com/ps-cert

Test report no.: 64111260284901

Date, 2026-05-22

(Xujun WANG)

Page 1 of 2

This Attestation does not replace the regulatory EU Declaration of Conformity (DoC) and does not allow for CE marking. After preparation of the necessary documentation and establishing compliance to requirements of all applicable directives, the manufacturer may sign a DoC and apply the CE marking. The DoC is issued under the sole responsibility of the manufacturer.



Attestation of Conformity

No. N8A 123678 0003 Rev. 00

Model(s): PF-A3, PF-A4, PF-A5, PF-A6

Brand: ANTECH

ANTECH

Parameters:

Rated Voltage:	230V
Rated Frequency:	50Hz
Rated Power:	2200W
Protection Class:	Class I
Degree of Protection:	IP X0

Tested according to: EN 61010-1:2010/A1:2019

Page 2 of 2

This Attestation does not replace the regulatory EU Declaration of Conformity (DoC) and does not allow for CE marking. After preparation of the necessary documentation and establishing compliance to requirements of all applicable directives, the manufacturer may sign a DoC and apply the CE marking. The DoC is issued under the sole responsibility of the manufacturer.