

CERTIFICATE

of Conformity Directive 2014/30/EU Electromagnetic Compatibility

Registration No.: AE 50739926 0001
Report No.: CN26JM70 004
Holder: Qingdao Antech Scientific Co., Ltd.
No. 182 Haier Road, Laoshan District,
Qingdao
266100 Shandong
P.R. China

Product: Laboratory Equipment
(Laboratory Refrigerator)

Type designation listed on the next page

This certificate of conformity is based on an evaluation of a sample of the above mentioned product. This is to certify that the tested sample is in conformity with all provisions of Annex I of Council Directive 2014/30/EU. This certificate does not imply assessment of the production of the product and does not permit the use of a TÜV Rheinland mark of conformity. The holder of the certificate is authorized to use this certificate in connection with the EC declaration of conformity according to the a.m. Directive.
This is not an EU-Type Examination Certificate.

Date: 2026-07-28

Certification Body



Ying Xie



TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg

CERTIFICATE

of Conformity Directive 2014/30/EU Electromagnetic Compatibility

Registration No.: AE 50739926 0001
Product: Laboratory Equipment
(Laboratory Refrigerator)

Tested according to: EN IEC 61326-1:2021

Identification: Type Designation

- 1) MPR-xxxxy
(xxx=22, 30, 40, 60, 110, 140;
y=blank or S)
MPR-xxxE
(xxx=60, 110, 140)
- 2) MPR-xxxxy
(xxx=40, 60, 110, 140; yy=A or SA)
MPR-xxxEA
(xxx=60, 110, 140)
- 3) MPR-xxxxy
(xxx=170, 200, 210, 260, 280, 300, 350,
410, 450; y=blank or S)
MPR-xxxE
(xxx=170, 190, 200, 260, 280, 300, 350,
410, 420)
- 4) MPR-xxxxxy
(xxxx=550, 650, 770, 900, 1070, 1200, 1240,
1400, 1420; yy=blank, S, B or P)
MPR-xxxxE
(xxxx=520, 620, 740, 870, 1040, 1200, 1240,
1360, 1420)

Serial No.: Engineering sample

Remark: Refer to test report CN26JM70 004 for details.



TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg