



Product Service

CERTIFICATE

No. Z1 089832 0018 Rev. 00

Holder of Certificate: SUZHOU ANTAI AIRTECH CO., LTD

No.2 Weixin Road, SIP
215122 Suzhou, Jiangsu
PEOPLE'S REPUBLIC OF CHINA

Certification Mark:



Product: Laboratory Equipment
Biosafety Cabinet

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition, the certification holder must not transfer the certificate to third parties. This certificate is valid until the listed date, unless it is cancelled earlier. All applicable requirements of the Testing, Certification, Validation and Verification Regulations of TÜV SÜD Group have to be complied. For details see: www.tuvsud.com/ps-cert

Test report no.: 4842023609000

Valid until: 2029-06-03

Date, 2024-06-04


(Bin Li)

CERTIFICATE

No. Z1 089832 0018 Rev. 00

Model(s): BSC-0605IIA2-SE-CN, BSC-1005IIA2-SE-CN,
BSC-1305IIA2-SE-CN, BSC-1605IIA2-SE-CN,
BSC-1805IIA2-SE-CN

Parameters:

Rated voltage	220-240VAC
Rated frequency	50/60Hz
Rated power	0.81KVA (BSC-0605IIA2-SE-CN); 1.65KVA(BSC-1005IIA2-SE-CN); 1.75KVA (BSC-1305IIA2-SE-CN); 1.85KVA (BSC-1605IIA2-SE-CN, BSC-1805IIA2-SE-CN)
Protection class	I
Degree of protection	IP20

Tested according to: EN 61010-1:2010/A1:2019
EN 12469:2000